

 RBmedikal Electro Surgical Instruments	Sterile & Non-Sterile, Single Patient Use Only		Document #: TFRBI-09-03/01
	INSTRUCTION FOR USE PROBES		Revision #: 01
			Doc Date: 10-05-2021
		Page 1 of 1	

In surgery, a probe is a slender/flexible rod with a blunt end used to explore, for example, an opening to see where it goes.

INTENDED USE:

Surgical Probes are used to manipulate tissue and identify nodes and other biological materials during surgery.

DISPOSABLE PRODUCT:

Check the packing against any damage, pore and tear.

Read carefully expiry date and intended purpose including Sterile Methodology

TESTING AND INSPECTION:

Before use inspect for rust, pits, cracks, pores & dents before use.



WARNING:

- Don't use for second time
- Product material is conductive take away from electric source.
- Don't use without Sterilization
- Don't use wrong instrument. It can be creating problem in operation
- Our instruments are designed for use by competent surgeons, who have a good knowledge of their features and how the instruments should be used. Any other use can compromise the safety of the user and the patient. It is the responsibility of the surgeon to choose the most suitable instrument/size/shape for the surgical technique being performed, based on his experience and expertise. So this product must be used by an expert surgeon/user.

RECOMMENDED CLEANING AND DISINFECT METHOD:

If product is supplied non-disinfected then product should be cleaned/disinfected with a mixture of 70% IPA & 30% Distilled water. Otherwise if products are supplied cleaned/disinfected then there is no need to re-clean or re-disinfect these. Sterilize all instruments before use.

RECOMMENDED STERILIZATION METHOD:

Non-Sterile Instruments can be sterilized by using following methods:

EO Sterilization:

70% CO / 30% ETO
Concentration ETO: 600mg/L
Temperature: 50°C - 55°C
Exposure time: 6hours
Humidity: 60-85% RH minimum
Pressure: 50-60kpa

CONSEQUENCES OF REUSE THE SINGLE USE PRODUCT

Reuse of single use instruments can cause cross contamination. It contains the risk of biological and chemical hazards. Re-sterilization of used single use instruments can cause the deterioration of instruments material as well.

WARRANTY:

This product is guaranteed to be free from defects in material or workmanship. The warranty is null and void should product damage occur as a result of mishandling or misuse. Care must be taken in the use and reprocessing of this product

SHELF LIFE:

3 years for Non-Sterile Packaging
3 years for Sterile Packaging

OPERATING CONDITIONS:

Follow your operating theater's SOP/Laws (These products are not for use in special temperatures.

STORAGE & TRANSPORT CONDITIONS:

Ambient temperature range: -05°C to +30°C
Relative humidity range: 40% to 70%
Atmospheric pressure range: 700 hPa to 1060 hPa

RECYCLE:

When reprocessing medical devices please follow your local government Health & Safety procedures

 RB MEDIKAL INSTRUMENTS S.I.E, Fateh Garh, Near Umer Town, Sialkot-51310, Pakistan Ph: +92 52 3552898 WhatsApp: +92-333-8619610 Email: rbmedikal@brain.net.pk rbmedikal@gmail.com Web: www.rbmedikal.com		
EC REP CMC MEDICAL DEVICES & DRUGS SL C/ Horacio Lengo, 18. 29006 Málaga-Spain	CH REP CMC MEDICAL DEVICES GMBH Rigistrasse 3, 6300 Zug, Switzerland Email: chrep@cmcmedicaldevices.com Ph: +41 415 620 395	Notified Body: SGS Belgium NV SGS House, Noorderlaan 87, 2030 Antwerpen, Belgium
 		

Explanation symbols used on labelling:

- | | | | |
|---|------------------------------------|---|-------------------------|
|  | Consult instructions for use |  | Keep away from sunlight |
|  | For single use only |  | Keep dry |
|  | Do not re-sterilise |  | Manufacturer |
|  | Do not use if packaging is damaged |  | Batch code |
|  | Used by/expiry date | | |

 Catalogue No.

 Sterilised by ETO
(for sterile delivered product only)

 European Union Authorised Representative

 Medical device

Signature:
Abdul Rehman
(CEO)